


| REVISIONS | | | |
|-----------|-------------|------|----------|
| REV | DESCRIPTION | DATE | APPROVED |
| | | | |



- NOTES:
- ALL DIMENSIONS ARE IN MM.
 - SEAL TEMP: SHOULD BE MAINTAINED BELOW 350°C
 - THE DIMENSIONS AND SPECIFICATIONS OF PRODUCT LISTED IN THIS SHEET ARE SUBJECT TO CHANGE WITHOUT NOTICE FOR IMPROVEMENT.

| ITEM | QTY | DESCRIPTION | CODE IDENT | PART NUMBER | REMARKS |
|---|-----|----------------------------|------------|--|---------|
| LIST OF MATERIALS | | | | | |
| UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES | | CONTRACT NO. | | USHIO AMERICA, INC. 5440 Cerritos Avenue Cypress, CA 90630 | |
| DECIMALS .X ± - | | APPROVALS | | | |
| .XX ± | | DATE | | | |
| .XXX ± | | DRAWN R. AMPARO 03/29/00 | | | |
| ANGLES ± | | CHKD D.GILCZYNSKI 03/29/00 | | TITLE | |
| DRILLED HOLES PER AND 10387 REMOVE BURRS AND SHARP EDGES ALL DIMENSIONS AFTER PLATING | | ENGR | | SM-8024 SCIENTIFIC MEDICAL LAMP | |
| MATERIAL | | MFG | | | |
| COMPOSITE | | QUAL | | | |
| FINISH | | DSGN ACTVY | | | |
| APPLICATION | | GOVT ACTVY | | SIZE B | |
| | | | | CODE IDENT NO. | |
| | | | | DWG NO. 8000271 | |
| | | | | SCALE NONE | |
| | | | | WEIGHT | |
| | | | | SHEET 1 OF 1 | |

| CHARACTERISTICS | | | |
|-----------------|---------|-------------|---------|
| VOLTAGE | 12 V | | |
| WATTAGE | 40 W | | |
| FILAMENT TYPE | C-6 | | |
| AVERAGE LIFE | 100 HRS | NEXT ASSY | USED ON |
| BASE TYPE | BA20D | APPLICATION | |

4 3 1

D

C

B

A

CAD